



Product Service

Attestation of Conformity

Galvanic Applied Sciences USA INC.

**41 Wellman St
Lowell, MA 01851
United States**

Item Identification: Spectrophotometer

Model(s): FPA 4000 Photometer; FPA 4100 Photometer

Item Description: Parameters

Width:	25.0cm
Depth:	25.0cm
Height:	67.0cm
Weight:	28.4Kg
Rated Input Voltage:	12-28Vdc 52W or 100-240V~
Rated Frequency:	50/60Hz
Rated Input Power:	100 W
Protection Class:	I

The equipment complies with the principal protection requirements of EN 61010-1: 2001 (2nd Edition) based on a voluntary test.

This attestation applies only to the particular sample of the product and its technical documentation provided for testing. The test results and all standards used are listed in.

Test report no. 090-1006814-000

Test standard: EN 61010-1: 2001 (2nd Edition)

This attestation does not permit the use of a TÜV PRODUCT SERVICE certification mark.

Department: Product Safety Services, Peabody MA
Date: 2010-09-17



TÜV®

ATTESTATION