



**QPS Evaluation Services Inc**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

File <b>LR1406</b>
-----------------------

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 3 CERTIFICATION SYSTEM)**

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Road SE Calgary, AB T2H 0W3
Project Number	LR1406-2
Product	Analytes in Water Analyzer
Model Number	AccuColor or AccuSense
Ratings/	100-240 Vac, 50/60 Hz, 310W
Markings	Class I, Division 2, Groups BCD, T3 Ta=-20°C to +60°C
Applicable Standards	ISA 12.12.01:2015 CSA C22.2 No.213-16 CSA C22.2 No. 61010-1:12 UL 61010-1 3rd ed.
Factory/Manufacturing Location	Same as address above

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: **Dave Adams, P.Eng.**  
 Manager, Hazardous Locations [Ex Equipment] Department

Signature:

Date: February 17, 2017