



Field Certification Certificate of Compliance

Certificate: 70155726

Master Contract: N/A

Project: 70155726

Date Issued: December 6, 2017

Issued to: Galvanic Applied Sciences Inc.
7000 Fisher Rd SE
Calgary, Alberta,

Sales Order: N/A

CANANDA

Attention: David Haydt

The products listed below are eligible to bear CSA Field Certification Labels, bearing the CSA Mark shown with adjacent indicators 'C' and 'US'.



Issued by: Tony DeSousa SCO/CET

PRODUCTS

CLASS 2258-02	PROCESS CONTROL EQUIPMENT-For Hazardous Locations
--------------------------------------	---

Class I, Division 2, Groups B, C and D, Temperature Code T3

ViscoSite viscosity transmitter VS7000
Electrical rating: 100-240 VAC @ 1.2A – 0.6A, 60 Hz

CSA Field Certification Label(s) issued: FB 559773