



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

Certificate Number
QPS 20ATEX0002X

CERTIFICATE OF COMPLIANCE
ATEX Essential Health & Safety Requirements
DIRECTIVE 2014/34/EU

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Road SE Calgary, Alberta, T2H 0W3
Project Number	ATX1406-3
Product	AccuLase Z2 analyzer
Model Number	AccuLase Z2
Markings	 II 3G Ex ec op is IIB+H2 T3 Gc, T _{amb} = -20°C to +50°C
Ratings	Input: 100-240 VAC, 1 Phase, 50/60 Hz, 230 Watts
Applicable Standards	For ATEX Directive 2014/34/EU Annex II and: EN IEC 60079-0 : 2018 EN IEC 60079-7:2015+A1:2018 EN 60079-28: 2015
Factory/Manufacturing Location	Galvanic Applied Sciences Inc. 7000 Fisher Road SE Calgary, Alberta, T2H 0W3
Specific Conditions of Use	<ol style="list-style-type: none"> 1. Read and understand the manual before working with the analyzer. 2. Only qualified personnel should install, operate, and maintain the analyzer. 3. INVISIBLE LASER RADIATION. CLASS 3b RADIATION PRODUCT. AVOID EXPOSURE TO THE BEAM (Locate Laser Detector Enclosure Door)

QPS Evaluation Services Inc., Unit #8, 81 Kelfield Street, Toronto, Ontario M9W 5A3, Canada, certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of the equipment intended for use in potentially explosive atmospheres given in Annex II to the Directive.

Issued By: Dave Adams P.Eng. Manager, Hazardous Locations Department

Signature: 

Date: March .8 2023