



**QPS Evaluation Services Inc**  
 Testing, Certification and Field Evaluation Body  
 Accredited in Canada, the USA, and Internationally

File <b>LR1406</b>
-----------------------

**CERTIFICATE OF COMPLIANCE**  
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Galvanic Applies Sciences Inc.
Address	7000 Fisher Rd. SE Calgary, AB T2H 0W3
Project Number	LR1406-5
Product	Air Demand Analyzer (Pit Gas Analyzer)
Model Number	943-TGXeNA (943-PGX-eNA)
Ratings	Class I. Division 2, Group C and D, T3 Tamb: -20°C to + 60°C. Type: 4, 4X, 12 and IP66
Applicable Standards	Input Voltage: 100 – 240 VAC, 50/60Hz, 1 Phase, 800 watts CSA C22.2 No. 61010-1-12 and UL 61010-1 3rd (as a Guide) CSA C22.2 No. 213-17 and UL 121201 2017. NFPA 496-2017
Factory/Manufacturing Location	Same as Applicant

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams, P. Eng.  
 Manager, Hazardous Locations [Ex Equipment] Department

Signature: 

Date: August 31, 2020