



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File LR1406

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Road SE Calgary, Alberta T2H 0W3
Project Number	LR1406-6R2
Product	Gas Chromatograph
Model Number	AccuChrome-D1-S AccuChrome-D1-C6 AccuChrome-D1-C9
Ratings	Class I, Division 1, Groups BCD T3 T _{amb} = -20°C to +60°C 100-240 VAC, 1 Phase, 50/60 Hz, 150 watts OR 24 VDC, 150 watts
Applicable Standards	CSA C22.2 No 61010-1-12 UL 6110-1 3 rd ed. CSA C22.2 No. 30: 2020 UL 1203 5 th ed.
Factory/Manufacturing Location	Same As Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams, P. Eng.
 Certification Manager, Hazardous Locations [Ex Equipment] Department

Signature:

Date: June 29, 2022